

**The LCV Current & Voltage Data Logger
measures external Current and Voltage,
simultaneously.**

Simultaneous Monitoring of Current and Voltage for:

- Compressor and Fan Run Cycles
- Troubleshooting Intermittent problems
- Verification of Power Conditions

SPECIFICATIONS

Sample Point Capacity:	10,750 points for Current & Voltage 21,500 points for Current OR Voltage
Sample Rate:	1 second to 18 hour
Alarms:	Over and under alarms for current and voltage Visual alarm indication
Current Range:	0 to 300 Amps AC
Current Accuracy:	+/- 5%
Current Resolution:	0.1 Amp AC
Voltage Range:	0 to 500 Volts AC
Voltage Accuracy:	+/- 1 V AC
Voltage Resolution:	0.1 V AC
Time Accuracy:	+/-100ppm @75°F
Operating Range:	
Temperature:	32°F to 140°F (0°C to 60°C) using Alkaline Battery. -40°F to 150°F (-40°C to 65°C) using Lithium Battery.
Humidity:	0% to 95%, Non-condensing
Storage Temperature:	-40°F to 170°F (-40°C to 77°C)
Calibration:	Single point offset calibration available through software. Current and Voltage are factory calibrated.
Tamper proof	
Dimensions:	3" x 2.5" x 1.1" (7.6cm x 6.4cm x 2.8cm)
Weight:	2.5 Oz (71g)
Power Source:	9V Alkaline or Lithium battery.
Battery Life (average use):	20 Months using Alkaline Battery 40 Months using Lithium Battery
1 year warranty	

PC Software operates with all LOGiT loggers (p/n LLS not included)



LCV

LOGiT

**Current & Voltage
Data Logger**

- One (1) Current Measurement
- One (1) Voltage Measurement

SEALED UNIT PARTS CO., INC.
PO BOX 21, 2230 LANDMARK PLACE
ALLENWOOD, NJ 08720 USA
www.supco.com • info@supco.com

LCV: AC Current, AC Voltage Logger

The LCV data logger is a precision instrument for recording AC current and AC voltage. AC current is sampled using the supplied clamp probe. The logger can store up to 21,500 data points, when it is setup to record either AC current or AC voltage (but not both). The logger can store up to 10,750 points, when it is setup to record both AC current and AC voltage. Each logger is factory calibrated. This means that the clamp probes are not interchangeable.

Connecting the LCV to Electrical Equipment

Before connecting to any equipment follow these precautions:

- Make sure that the wires are securely attached to the connector and that no uninsulated portion of the wire is visible after the connection is made.
- Make sure that there is no strain that could pull the leads out of the LCV connector.
- Make sure that the leads will not interfere with any moving pieces of equipment and will not pull loose any wires or other connections.

To measure the voltage, connect the two voltage clip leads to the points at which the voltage is being measured.

To measure the current, place the clamp probe around the wire in which the current is being measured.

CAUTION!

*Electrical connection to any type of operating equipment is **Dangerous!** The LCV leads and current probe should only be connected when it has been determined that the power is off and the equipment is safe to touch. It is recommended that you always remove the LCV voltage leads before connecting the LCV to your computer!*

LCV Specifications

Sample Point Capacity	10,750 data points when AC current and AC voltage are recorded together. 21,500 points when recording only AC current or only AC voltage.
Alarms	Over and under alarms for AC current and AC voltage.
Calibration	AC current and AC voltage are factory calibrated. User single point offset calibration is available through software for both AC current and AC voltage.
Operating Temperature	32°F to 140°F (0°C to 60°C) using Alkaline Battery. -40°F to 150°F (-40°C to 65°C) using Lithium Battery.
Time Accuracy	+/-100ppm @75°F
Relative Humidity Range	0% to 95%RH non condensing
Current Range	0 to 300 Amp AC
Current Accuracy	+/- 5%
Current Resolution	0.1 Amp AC
Voltage Range	0 to 500 Volts AC
Voltage Accuracy	+/- 1 VAC
Voltage Resolution	0.1 VAC
Dimensions	3" x 2.5" x 1.1" (7.6cm x 6.4cm x 2.8cm)
Weight	2.5 Oz (71g)
Power Source	9V Alkaline or Lithium battery.
Battery Life (average use)	18 Months using Alkaline Battery. 36 Months using Lithium Battery.

LOGiT Common Specifications

Sampling Interval	User selectable from 1 second to 18 hours, in 1 second increments.
Interface	COM (RS232) serial port. Software supports COM1 through COM8. Optional USB interface.
Alarms	Visual alarms with user selectable low and high thresholds for every channel.
Alarm Memory	Up to 50 alarm events. Alarm start time and alarm duration is stored.
Data Security	Locking sliding door prevents physical access to the logger data. Unique factory set logger id. Data collection session counter.
Compatibility Minimum System Requirements	Windows 95 and above. 8MB of RAM, 2M of disk space, 800 x 600 resolution, 1 free COM port or a free USB port (optional USB interface is required).
Recording Start	On door close, immediate after setup, or at a user specified data and time.
Recording Mode	Stop on memory full, or continuous recording with memory rollover.
Real Time Status	Visual indication through the status LED. User selectable status display frequency and LED brightness. Software real time status.
User Calibration	Password protected calibration for each channel.
User Message	Up to 30 characters.
Logger Channels	User selectable. Disabling channels increases the storage capacity of the other channels
Units	US and Metric. User selectable units.
Data Format	Custom format exportable to text or Excel spreadsheet.
Real Time Monitoring	Logger can be permanently connected to the computer and used as a real time, multichannel sensor with logging capability. Battery does not drain when logger is connected to the computer.